



America

CERTIFICATE

No. U8 098506 0007 Rev. 00

Holder of Certificate: **Pavemetrics Systems Inc.**
150 boul. René-Lévesque Est, suite 1820
Quebec QC G1R 5B1
CANADA

Production Facility(ies): 106278

Certification Mark:



Product: **Electrical equ. for measurement, control and laboratory use**
Measurement equipment

Model(s): LCMS-2

Parameters: Rated Input Voltage: 12-24 Vdc
Rated Input Current: 12.5 A
Protection Class: III

Tested according to: CAN/CSA-C22.2 No. 61010-1:2012/A1:2018-11
UL 61010-1:2012/R:2019-07

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 7169008600-000

Date, 2020-12-17

(Vivek Chandna)